



NATIONAL TYPE EVALUATION PROGRAM

# Certificate of Conformance

for Weighing and Measuring Devices

**For:**  
Load Cell  
Double Ended Shear Beam, Direction Indicated by Arrow  
Model: 102BH Series  
Multiple Cell, Class III L  $n_{max}$ : 10 000  
Capacity: 25 000 to 125 000 lb  
Accuracy Class: III L

**Submitted By:**  
Anyload Transducer Co., Ltd  
Unit 102, 6994 Greenwood Street  
Burnaby, BC V5A1X8  
Canada  
Tel: 604-420-2130  
Fax: 866-612-9088  
Contact: Gary Gui  
Email: [gary.gui@load-cell.com](mailto:gary.gui@load-cell.com)  
Web site: [www.anyload.ca](http://www.anyload.ca)

**Standard Features and Options**

The specific load cell capacities,  $v_{min}$  values, and minimum dead loads covered by this Certificate are listed in the table below.

**Standard Features:**

- Nominal Output: 3.0 mV/V
- Alloy Steel
- 4 Wire Design


Model	Capacity	$V_{min}$ Class III L Multiple Cell, $n_{max}$ = 10 000	Minimum Dead Load
102BH*	25 000 lb	1.3 lb	0 lb
102BH	40 000 lb	2.0 lb	0 lb
102BH	50 000 lb	2.6 lb	0 lb
102BH	60 000 lb	3.1 lb	0 lb
102BH	75 000 lb	3.9 lb	0 lb
102BH	100 000 lb	5.2 lb	0 lb
102BH	125 000 lb	6.5 lb	0 lb

\*Load Cell Tested

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

  
Tim Tyson  
Chairman, NCWM, Inc.

  
Randy Jennings  
Chairman, National Type Evaluation Program Committee  
Issued: March 22, 2011

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



## Anyload Transducer Co., Ltd

### Load Cell / 102BH Series

**Application:** The load cells may be used in Class III L scales for multiple cell applications consistent with the model designations, number of scale divisions, and parameters specified in this certificate. Load cells of a given accuracy class may be used in applications with lower accuracy class requirements provided the number of scale divisions, the  $v_{\min}$  value, and temperature range are suitable for the application. The manufacturer may market the load cell with fewer divisions ( $n_{\max}$ ) and with greater  $v_{\min}$  values than those listed on the certificate. However, the load cells must be marked with the appropriate  $n_{\max}$  and  $v_{\min}$  for which the load cell may be used.

**Identification:** A pressure sensitive identification label located on the cell, states manufacturer name, model number, serial number, rated capacity, class,  $v_{\min}$ , direction of loading and CC number. Other pertinent information will be specified on the Calibration Certificate accompanying the cell.

**Test Conditions:** A Model 102BH (25 000 lb capacity) load cell was tested by the NMI Certin B.V. at the Netherlands facility. Testing was conducted in accordance with the OIML DoMC Mutual Acceptance Arrangement, signed by the NCWM as a utilizing participant for load cell testing. Testing was conducted using deadweights as the reference standard. The load cells were tested over a temperature range of -10 °C to 40 °C with tests run on each cell at each temperature. The temperature effect on zero was measured and a time dependence (creep) test was performed. The barometric pressure test to determine sensitivity of the load cell design to changes in barometric pressure was conducted. The data were analyzed for multiple load cell applications. OIML R60 selection criteria was used to determine cells tested.

**Evaluated By:** A.C Pauwels, R. Scholten (NMI)

**Type Evaluation Criteria Used:** NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 2011. NCWM, Publication 14: Weighing Devices, 2011.

**Conclusion:** The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

**Information Reviewed By:** J. Truex (NCWM)

**Example of Device:**

