



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Load Cell
 Compression
 Model: 108XA Series
 n_{max} : 4 000, Class III, Single Cell
 Capacity: 200 to 2000 kg
 Accuracy Class: III

Submitted By:

Anyload Transducer Co., Ltd
 Unit 102, 6994 Greenwood Street
 Burnaby, BC V5A 1X8
 Canada
 Tel: 604-420-2130
 Fax: 866-612-9088
 Contact: Gary Gui
 Email: gary.gui@load-cell.com
 Web site: www.anyload.net

Standard Features and Options


- Model 108XA, where the X in the model designation may be C, J, M or U
- The specific load cell capacities, v_{min} values, and minimum dead loads covered by this Certificate are listed in the table below.
- Nominal output: 2.0 mV/V
- Aluminum
- 4 Wire Design
- Minimum Dead Load: 0 kg

Models	Capacity	v_{min} Class III Single Cell, n = 4 000
108CA, 108JA, 108MA and 108UA	200 kg	0.027 kg
	250 kg	0.033 kg
	300 kg	0.04 kg
	500 kg*	0.067 kg
	600 kg	0.08 kg
	700 kg	0.09 kg
	1 000 kg	0.13 kg
	1 500 kg	0.20 kg
	2 000 kg	0.27 kg


* load cell tested

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.



Kurt Floren
 Chairman, NCWM, Inc.



Jim Tyson
 Chairman, National Type Evaluation Program Committee
 Issued: March 20, 2012

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Anyload Transducers Co., Ltd

Load Cell / 108XA Series

Application: The load cells may be used in Class III scales for single cell applications consistent with the model designations, number of scale divisions, and parameters specified in this certificate. Load cells of a given accuracy class may be used in applications with lower accuracy class requirements provided the number of scale divisions, the v_{\min} value, and temperature range are suitable for the application. The manufacturer may market the load cell with fewer divisions (n_{\max}) and with greater v_{\min} values than those listed on the certificate. However, the load cells must be marked with the appropriate n_{\max} and v_{\min} for which the load cell may be used.

Identification: A pressure sensitive identification label located on the cell, states manufacturer name, model, serial number, rated capacity, class, n_{\max} and v_{\min} . Other pertinent information will be specified on the Calibration Certificate accompanying the cell.

Test Conditions: A Model 108JA, 500 kg capacity load cell was tested by the NMI Certain B.V. at The Netherlands facility. Testing was conducted in accordance with the OIML DoMC Mutual Acceptance Arrangement, signed by the NCWM as a utilizing participant for load cell testing. Testing was conducted using deadweights as the reference standard. The load cells were tested over a temperature range of -10°C to 40°C with tests run on each cell at each temperature. The temperature effect on zero was measured and a time dependence (creep) test was performed. The barometric pressure test to determine sensitivity of the load cell design to changes in barometric pressure was conducted. The data were analyzed for single load cell applications. OIML R60 selection criteria were used to determine cells tested.

Evaluated By: A.C. Pauwels, R. Scholten (NMI)

Type Evaluation Criteria Used: NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 2012. NCWM, Publication 14: Weighing Devices, 2012.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: J. Truex (NCWM)

Example of Device:

