



OIML CERTIFICATE OF CONFORMITY

To YOUNGZON TRANSDUCER (HANGZHOU) CO., LTD

This certificate attests the conformity of the type of the Load
Cell YZ563 with the requirements
of the Recommendation R60(2000) of the International
Organization of Legal Metrology(OIML).

The issuing authority
Office of OIML Affairs
中国秘书处
March 23, 2004

The OIML member

March 23, 2004

Member State
P.R.China

OIML Certificate No.
R60/2000-CN-04.01

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine
Address: No.9 Madiandonglu, Haidian District, Beijing 100088 China

Applicant

Name: YOUNGZON TRANSDUCER (HANGZHOU) CO., LTD
Address: No.1, Building Yile Industry Garden, No.75 Wenhua Rd., Hangzhou, 310012, P. R. China

Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:
Address:

Identification of the certified type

Model designation	YZ563				
Accuracy class	C3				
Maximum capacity (E_{max})	500kg	1t	2t	5t	10t
Minimum load cell verification interval (V_{min})	150g	300g	600g	1.5kg	3kg
Minimum dead load (E_{min})	10kg	10kg	10kg	20kg	50kg
Maximum number of load cell intervals (n_{max})	3000				
Apportionment factor (P_{LC})	0.7				
Safe overload	150% E_{max} (kg)				
Operating temperature range	-10°C - +40°C				
Humidity condition	CH				

The numbers of the associated test reports and the included pages:
2003-WJ-0002 2003-WJ-0006 18pages

This certificate attests the conformity of the above-mentioned type (represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 60

Edition: 2000

For accuracy class (if applicable): C3

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test reports.

The issuing authority

Office of OIML Affairs

中国秘书处

March 23, 2004

The OIML member

王春平

March 23, 2004

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.