

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMi Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Anyload Transducer Co. Ltd.
Address: 7228 Winston Street,
#18, Burnaby, BC, V5A 2G9,
Canada

Manufacturer of the certified type

Name: Anyload Youngzon Transducer (Hangzhou) Co, Ltd.
Address: No. 160, South No. 11 Street,
Hangzhou Economic & Technological Development Zone, Zhejiang
China, 310018

Identification of certified type

A Single point load cell
Type : 108JA...
Fraction : $P_i = 0.7$
Temperature range -10 °C / 40 °C

For specifications, see page 2.

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the test certificate and the description with number TC7692 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60
Edition 2000 (E)
for accuracy class C

OIML Member state

The Netherlands

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R60/2000-NL1-10.03 that includes 63 pages.

The Issuing Authority NL1
NMI Certin, 24 February 2010

C. Oosterman
Head Certification Board

*
* *

Load cell specifications:

Type	108JA
Maximum capacity (E_{max})	500 kg up to and including 2500 kg
Humidity Class	CH
Accuracy Class	C
Maximum number of load cell intervals (n)	4000
Ratio of minimum LC Verification interval $Y = E_{max} / V_{min}$	4000
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	7500

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.